

Mechanical Ventilation: Settings and Basic Modes

Mechanical ventilation is a lifesaving tool utilized in the intensive care unit and emergency department, during surgical procedures requiring general anesthesia, and in long-term acute care settings to assist patients who require additional respiratory support. Invasive mechanical ventilation requires an endotracheal tube that is connected to a ventilator, whereas non-invasive ventilation delivers oxygen through a face mask. Both types can be used for acute or chronic hypoxemic or hypercarbic respiratory failure, which is defined as insufficient oxygenation, insufficient alveolar ventilation, or both. Invasive mechanical ventilation is also useful for those who require airway protection to reduce the risk of aspiration (e.g., depressed mental status from an overdose or brain injury) as well as in the operating room during procedures requiring general anesthesia/deep sedation.

Mechanical ventilation aids in improving gas exchange, decreases work of breathing and supports the patient during treatment of the underlying condition. Ventilator settings integrate volume, flow, pressure, and time. Use this handy reference guide to help you safely manage and optimize oxygenation and ventilation goals for your patients with invasive/non-invasive mechanical ventilation.

MECHANICAL VENTILATION: IMPORTANT TERMS TO KNOW	
Term	Clinical Considerations
Fraction of inspired oxygen (FiO₂)	<ul style="list-style-type: none"> • Concentration of oxygen in the inspired air • Use the lowest FiO₂ that achieves the targeted oxygenation • Avoid prolonged FiO₂ greater than 0.60, as this may cause oxygen toxicity and worsen hypoxemia
Frequency (f) or respiratory rate (RR) (10-20 breaths/min)	<ul style="list-style-type: none"> • Set number of ventilator breaths per minute • Actual RR includes the spontaneous breaths taken by the patient • Hypoventilation may cause respiratory acidosis; hyperventilation may cause respiratory alkalosis
Trigger	Breaths can be triggered by: <ul style="list-style-type: none"> • Timer (ventilator-initiated breaths); occur at the set respiratory rate or frequency (f) • Patient effort (patient-initiated breaths); occur when the patient causes sufficient change in either the pressure or flow in the circuit
Target	<ul style="list-style-type: none"> • Flow of air into the lung can target a predetermined flow rate (e.g., peak inspiratory flow rate) or pressure limit
Termination	<ul style="list-style-type: none"> • Signal for a ventilator to end inspiration • May be related to volume (e.g., tidal volume), time (e.g., predetermined duration of inspiration), or flow (decrease in inspiratory flow to a percentage of peak value)

MECHANICAL VENTILATION: IMPORTANT TERMS TO KNOW	
Term	Clinical Considerations
Tidal volume (V_T) (4-8 mL/kg of ideal body weight [IBW] to prevent barotrauma)	<ul style="list-style-type: none"> • Volume of gas exchanged with each breath • A lower V_T, or lung protective ventilation, is indicated in patients with stiff, non-compliant lungs, such as in acute respiratory distress syndrome (ARDS) • Higher V_T may cause tachycardia, decreased blood pressure and lung injury or barotrauma
Minute ventilation (V_E) (5-10 L/minute)	<ul style="list-style-type: none"> • Volume of gas exchanged per minute • $V_E = RR \times V_T$
Inspiratory: expiratory (I:E) ratio	<ul style="list-style-type: none"> • Normal: longer expiratory phase than inspiratory phase (1:2, 1:3) • Inverse ratios provide a longer inspiratory phase (1:1, 2:1, 3:1, 4:1) • Reduced I:E allows more time for exhalation and reduces breath stacking; used for patients who have obstructive airway disease with acute respiratory acidosis (e.g., acute asthma exacerbation)
Peak inspiratory pressure (PIP)	<ul style="list-style-type: none"> • Highest proximal airway pressure reached during inspiration • Target PIP is less than 35 cm H₂O • Low PIP may result in hypoventilation; high PIP may cause lung damage or barotrauma
Peak flow rate	<ul style="list-style-type: none"> • Maximum flow delivered by the ventilator during inspiration
Plateau pressure (Pplat) (Target Pplat less than 30)	<ul style="list-style-type: none"> • Reflects pulmonary compliance and is measured by applying a brief inspiratory pause after ventilation • Assess Pplat with peak inspiratory pressure (PIP): <ul style="list-style-type: none"> ○ A high PIP with normal Pplat = increased resistance to flow (e.g., endotracheal tube obstruction or bronchospasm) ○ High PIP and high Pplat = decreased lung compliance (e.g., interstitial pulmonary fibrosis, pneumonia, ARDS, pulmonary edema)
Positive end expiratory pressure (PEEP) (3-10 cm H ₂ O)	<ul style="list-style-type: none"> • Pressure remaining in the lungs at end expiration • Used to keep alveoli open and “recruit” more alveoli to improve oxygenation for patients • High levels may cause barotrauma, increased intracranial pressure and decreased cardiac output

MECHANICAL VENTILATION: IMPORTANT TERMS TO KNOW	
Term	Clinical Considerations
Driving Pressure (ΔP) (target 15 cm H ₂ O or less)	<ul style="list-style-type: none"> The pressure required to open the alveoli $\Delta P = P_{plat} - PEEP$ High driving pressure (greater than 15) indicates poor lung condition with decreased lung compliance.
Pressure support (PS) (8-20 cm H ₂ O)	<ul style="list-style-type: none"> Provides additional pressure during inspiration to ensure a larger V_T with minimal patient effort Used to help overcome the work of breathing through ventilator tubing Can be used as weaning method while working towards ventilator liberation

Modes of Mechanical Ventilation

MODES OF MECHANICAL VENTILATION	
MODE	DESCRIPTION
VOLUME-LIMITED VENTILATION Clinician sets the peak flow rate, flow pattern, V _T , RR, PEEP, and FiO ₂ ; inspiration ends once the set inspiratory time has elapsed.	
Controlled mechanical ventilation (CMV)	<ul style="list-style-type: none"> V_E is determined entirely by the set RR and V_T. No patient effort required.
Assist control (AC)	<ul style="list-style-type: none"> Minimal V_E is determined by setting the RR and V_T. Patient can increase the V_E by triggering additional breaths. Each patient-initiated breath receives the set V_T from the ventilator.
Pressure regulated volume control (PRVC) (also called volume control plus [VC+])	<ul style="list-style-type: none"> V_T is set and the applied airway pressure changes to attain the target V_T. The initial applied inspiratory pressure is determined by the change in pressure required by the previous breath to attain the V_T.
Intermittent mandatory ventilation (IMV)	<ul style="list-style-type: none"> Minimal V_E is determined by setting the RR and V_T; patient can increase the V_E by spontaneous breathing, rather than patient-initiated ventilator breaths.
Synchronized intermittent mandatory ventilation (SIMV)	<ul style="list-style-type: none"> A variation of IMV; ventilator breaths are synchronized with patient inspiratory effort; can be used to titrate the level of ventilatory support.

MODES OF MECHANICAL VENTILATION	
MODE	DESCRIPTION
<p>PRESSURE-LIMITED VENTILATION Clinician sets the inspiratory pressure level, I:E ratio, RR, PEEP, and FiO₂; inspiration ends after delivery of the set inspiratory pressure.</p>	
<p>Pressure-limited CMV (also called pressure control ventilation)</p>	<ul style="list-style-type: none"> • V_E is determined entirely by the set RR and inspiratory pressure. • Patient does not initiate additional V_E above that set on the ventilator.
<p>Pressure-limited AC</p>	<ul style="list-style-type: none"> • Set RR and inspiratory pressure determine the minimum V_E. • The patient can increase the V_E by triggering additional ventilator-assisted, pressure-limited breaths.
<p>Pressure-limited IMV or pressure-limited SIMV</p>	<ul style="list-style-type: none"> • The set RR and inspiratory pressure determine the minimum V_E. • The patient can increase the V_E by initiating spontaneous breaths.
<p>Pressure support</p>	<ul style="list-style-type: none"> • Flow-limited mode that delivers inspiratory pressure until the inspiratory flow decreases to a predetermined percentage of its peak value. • Clinician sets the pressure support (inspiratory pressure), applied PEEP, and FiO₂. The patient must trigger each breath because there is no set RR. The V_T, RR, and V_E are dependent on multiple factors, including ventilator settings and patient-related variables. • The work of breathing is inversely proportional to the pressure support level and the inspiratory flow rate.
<p>Continuous positive airway pressure (CPAP)</p>	<ul style="list-style-type: none"> • Delivery of a continuous level of positive airway pressure. • Functionally similar to PEEP. • The ventilator does not cycle during CPAP and no additional pressure above the level of CPAP is provided; patients must initiate all breaths.
<p>Bilevel positive airway pressure (BiPAP)</p>	<ul style="list-style-type: none"> • Noninvasive positive pressure ventilation (NPPV) that delivers a preset inspiratory positive airway pressure (IPAP) and expiratory positive airway pressure (EPAP). The V_T correlates with the difference between the IPAP and the EPAP.

MODES OF MECHANICAL VENTILATION	
MODE	DESCRIPTION
Airway pressure release ventilation (APRV)	<ul style="list-style-type: none"> • High continuous positive airway pressure (P high) is delivered for a long duration (T high) and then falls to a lower pressure (P low) for a shorter duration (T low). • Alveolar recruitment is maximized by the high continuous positive airway pressure. • V_T is related to both the driving pressure and the compliance. • Must allow for permissive hypercapnia
Inverse ratio ventilation (IRV)	<ul style="list-style-type: none"> • The inspiratory time exceeds the expiratory time. • Strategy employed during volume-limited or pressure-limited mechanical ventilation to increase the mean airway pressure and potentially improve oxygenation.

References:

Hyzy, R. & MsParron, J. (2025, June 24). Overview of initiating mechanical ventilation in adults in the intensive care unit. *UpToDate*. <https://www.uptodate.com/contents/overview-of-initiating-invasive-mechanical-ventilation-in-adults-in-the-intensive-care-unit>

Hyzy, R. & Jia, S. (2025, December 10). Modes of mechanical ventilation. *UpToDate*. <https://www.uptodate.com/contents/modes-of-mechanical-ventilation>

Patel, B. (2024). Overview of Mechanical Ventilation. *Merck Manual Professional Edition*. <https://www.merckmanuals.com/professional/critical-care-medicine/respiratory-failure-and-mechanical-ventilation/overview-of-mechanical-ventilation>

Zaidi, S. F., Shaikh, A., Khan, D. A., Surani, S., & Ratnani, I. (2024). Driving pressure in mechanical ventilation: A review. *World journal of critical care medicine*, 13(1), 88385. <https://doi.org/10.5492/wjccm.v13.i1.88385>